

Portable microfluidic chip for detection of *Escherichia coli* in produce and blood

ShuQi Wang^{1*}

Fatih Inci^{1*}

Tafadzwa L Chaunzwa¹

Ajay Ramanujam¹

Aishwarya Vasudevan¹

Sathya Subramanian¹

Alexander Chi Fai Ip¹

Banupriya Sridharan¹

Umut Atakan Gurkan¹

Utkan Demirci^{1,2}

¹Bio-Acoustic-MEMS in Medicine (BAMM) Laboratory, Center for Biomedical Engineering, Department of Medicine, Brigham and Women's Hospital, Harvard Medical School, Boston, MA 02139, ²Harvard-MIT Health Sciences and Technology, Cambridge, MA, USA

*These authors contributed equally to this work

Correspondence: Utkan Demirci
Health Sciences and Technology,
Harvard-MIT Health Sciences
and Technology, Bio-Acoustic-MEMS
in Medicine (BAMM) Laboratory,
65 Landsdowne St, No 267,
02139 Cambridge, MA, USA
Tel +1 650 9069227
Email udemirci@rics.bwh.harvard.edu

Abstract: Pathogenic agents can lead to severe clinical outcomes such as food poisoning, infection of open wounds, particularly in burn injuries and sepsis. Rapid detection of these pathogens can monitor these infections in a timely manner improving clinical outcomes. Conventional bacterial detection methods, such as agar plate culture or polymerase chain reaction, are time-consuming and dependent on complex and expensive instruments, which are not suitable for point-of-care (POC) settings. Therefore, there is an unmet need to develop a simple, rapid method for detection of pathogens such as *Escherichia coli*. Here, we present an immunobased microchip technology that can rapidly detect and quantify bacterial presence in various sources including physiologically relevant buffer solution (phosphate buffered saline [PBS]), blood, milk, and spinach. The microchip showed reliable capture of *E. coli* in PBS with an efficiency of $71.8\% \pm 5\%$ at concentrations ranging from 50 to 4,000 CFUs/mL via lipopoly-saccharide binding protein. The limits of detection of the microchip for PBS, blood, milk, and spinach samples were 50, 50, 50, and 500 CFUs/mL, respectively. The presented technology can be broadly applied to other pathogens at the POC, enabling various applications including surveillance of food supply and monitoring of bacteriology in patients with burn wounds.

Keywords: *Escherichia coli*, microchip, sepsis, food safety, point-of-care

Introduction

Sepsis is a significant, life-threatening problem for military personnel and civilians alike. Clinicians typically make an initial diagnosis of infection with an intermittent examination observing changes in temperature, blood pressure, smell, and sight. Depending on the severity of the inflicted wound and infectious agent, this may cause multiple organ dysfunction, hypoperfusion, failure of body systems (including the kidneys, liver, lungs, and central nervous system), and ultimately death. For sepsis diagnosis, daily culture and/or molecular analysis is performed using a small volume of the patient's blood; usually less than 0.1% of entire blood volume (ie, approximately 5 mL). These detection methods are time-consuming and unreliable for identification of pathogens in up to 50% of septic patients. Following pathogen detection, the patient is treated with broad-spectrum antibiotics; this is not an ideal solution, because frequent use of broad-spectrum antibiotics also stimulates development of antibiotic-resistant bacteria. Antibiotic-resistant bacteria are a significant problem for injured military personnel and military treatment facilities. Currently, the sepsis mortality rate is 30%–50% among civilians and is higher in patients with military conditions including burns, trauma, and surgery.¹ Although sepsis biomarkers, inflammatory modulators, and new antibiotics have been identified in past sepsis research, improved identification of various targets has

had minimal impact on sepsis related morbidity and mortality. Therefore, there is a significant clinical need for new detection and identification technologies in this area. According to the Centers for Disease Control and Prevention, food-borne diseases cause approximately 76 million illnesses, 325,000 hospitalizations, and 5,000 deaths in US alone each year.² The US Department of Agriculture showed that medical costs, productivity losses, and costs of premature deaths caused by food-borne diseases are approximately \$6.9 billion per year.³ In 2011, pathogen-based produce contamination triggered a global concern with the outbreak of the foodborne toxin ‘Shiga’, which is produced by *Escherichia coli* (*E. coli*).⁴ On the other hand, sepsis is the tenth leading cause of death in the US,⁵ amounting to 24,179 cases in 49 US hospitals over a period of 7 years.⁶ As reported, *E. coli* can contaminate food source⁷ and cause sepsis in burn patients.^{8,9} Thus, effective *E. coli* detection would have a positive impact.

Currently, the gold standard detection method for bacteria is agar plate culture. However, this method is limited by the

culturing time and volume of sample required to determine the presence of pathogens (Figure 1). Due to the challenge of obtaining enough sample volume, agar plate cultures give false negative results at rates ranging from 7.2% to 21.2%.^{10,11} In addition, the process is complicated by the fact that clinical samples need to go through multiple post-cultural steps for analysis, including Giemsa staining and differentiation on MacConkey plates.¹² The whole process takes 48 to 72 hours.¹³ Although polymerase chain reaction (PCR) has high sensitivity and specificity,¹⁴ the need for a thermal cycler makes it unsuitable for point-of-care (POC) testing.¹⁵ Therein lies the niche for which microfluidic technologies are ideal; they have been employed to develop POC testing devices because of low manufacturing cost, reduced consumption of samples and reagents, and shortened assay time.^{15–20} However, existing microfluidic devices for bacterial detection, either based on PCR²¹ or enzyme-linked immunosorbent assay (ELISA), require multiple sample processing steps prior to detection.^{22,23} All of these methods suffer from challenges




Figure 1 Comparison of the conventional culture method and the microchip based *E. coli* detection. **(A)** Conventional procedure for bacteria detection in clinical facilities. Blood sample collection. (1) Blood samples are incubated in an automated blood culture system. (2) Pathogen or bacteria grown on agar plate are subject to Gram-staining for differentiation between Gram-positive and negative strains. (3) The sample is sub-cultured into a nutrient-rich agar plate for the identification of the species and to determine the bacterial concentration. **(B)** POC testing approach for rapid detection. Blood sample collection (spiked with GFP-expressing *E. coli* BL21 stock as a model microorganism). (1) The blood sample is analyzed in microchannels functionalized with *E. coli* antibodies. *E. coli* were specifically captured by antibodies on the microchannel surface. (2) Unbound *E. coli* are washed away with PBS using a syringe micropump. (3) GFP-expressing *E. coli* are imaged/counted under a fluorescence microscope.

Abbreviations: *E. coli*, *Escherichia coli*; GFP, green fluorescent protein; LBP, lipopolysaccharide binding protein; PBS, phosphate buffered saline; POC, point-of-care.

including culture time, need of high sample volumes and reagents, the requirement for preprocessing of samples, low accuracy of the pathogen detection, and high cost. Further, for the detection of rare bacteria, PCR and ELISA based methods require large initial sample volumes, and preprocessing of samples, and sample amplification. Thus, there is an unmet need to develop POC devices that can address these issues, and capture, isolate and detect bacteria from biologically complex samples such as blood and produce.

To address this unmet need, we developed a POC microchip for capture, isolation, and detection of *E. coli* in various samples such as physiological buffer solution (phosphate buffered saline [PBS]), blood, milk, and spinach in a simple and rapid manner. This microchip technology could be broadly used as a POC device for multiple applications to rapidly screen for bacteria contamination in blood and food samples, thus improving healthcare and food safety.

Materials and methods

Device fabrication

The microfluidic device was fabricated as previously reported.^{18,24,25} The device was designed with dimensions

of 22 mm × 60 mm with three parallel microchannels. To assemble this device, poly(methyl methacrylate) (PMMA) (1.5 mm thick; McMaster Carr, Atlanta, GA) and double-sided adhesive film (DSA) (50 μm thick; iTapestore, Scotch Plains, NJ) were cut using a laser cutter (Versa Laser™, Scottsdale, AZ). The PMMA base and a glass cover slip were then assembled via the DSA. In the assembled *E. coli* detection device, three microchannels (with dimensions of 50 mm × 4 mm × 50 μm in the DSA layer) were formed with an inlet and outlet (0.565 mm in diameter) at each end of the channels in the DSA layer. Before assembling the chip, glass cover was cleaned with ethanol using sonication. Then, it was washed with distilled water and dried under nitrogen gas. After cleaning steps, the glass cover was plasma treated for 60 seconds. Then, PMMA, DSA, and glass cover were assembled to form the complete microchip (Figure 2A).

Strains used in studies: genetically modified *E. coli*

To validate the surface chemistry, a genetically modified *E. coli* strain expressing emerald green fluorescent protein (EmGFP) was used. The *E. coli* strain BL21 Star™, and




Figure 2 Evaluation of two different surface chemistry methods for *E. coli* detection on chip. **(A)** Assembly of the microfluidic chip consisting of PMMA, DSA, and glass cover. Actual image of the assembled microchip containing food dye for visualization. **(B)** Two antibody immobilization mechanisms were employed, ie, Protein G and NeutrAvidin based surface chemistry. In the first method, biotinylated anti-LBP antibody was immobilized on the microchannel surface via NeutrAvidin. Then, LBP was immobilized on anti-LBP antibody. In the second method, CD14, anti-LPS, or anti-flagellin antibodies was immobilized on the microchannel surface via Protein G. Only CD14 immobilization was illustrated and similar steps were followed for anti-flagellin and anti-LPS. **(C)** Detection of GFP-tagged *E. coli* on-chip. To validate the *E. coli* capture process, and quantify the on-chip concentration and capture efficiency of *E. coli*, these cells were identified under brightfield (10× magnification) and fluorescence microscopy. (i) Image of the control experiment without *E. coli* at 10× magnification under a fluorescence microscope. (ii) Image of the capture of GFP-tagged *E. coli* at 10× magnification under a fluorescence microscope. (iii) Image of the capture of GFP-tagged *E. coli* at 100× magnification under a fluorescence microscope. (iv) Image of the captured GFP-tagged *E. coli* at 100× magnification under bright field.

Abbreviations: DSA, double-sided adhesive film; *E. coli*, *Escherichia coli*; GFP, green fluorescent protein; LBP, lipopolysaccharide binding protein; LPS, lipopolysaccharide; PBS, phosphate buffered saline; PMMA, poly(methyl methacrylate); POC, point-of-care; GMBS, N-(gamma-maleimidobutyryloxy) succinimide.

EmGFP-expressing plasmid were purchased from Invitrogen (Carlsbad, CA). According to the manufacturer's instructions, the pRSET/EmGFP plasmid, which confers ampicillin resistance and green fluorescence, (Invitrogen, catalog number V353-20) was transferred into BL21 Star™ by incubating at 41°C for 30 seconds and transferring onto ice. The transformed BL21 Star™ was then incubated at 37°C by shaking at 250 rpm for an hour in Super Optimal broth with catabolite repression medium. Following this step, BL21 Star™ was streaked onto Luria–Bertani (LB) agar plates containing 100 mg/mL of ampicillin, and then incubated at 37°C for 16 hours. An isolated *E. coli* colony was picked and inoculated in 5 mL of LB medium on another LB agar plate, with 100 mg/mL of ampicillin added. The *E. coli* culture was then incubated at 37°C with shaking at 250 rpm for 16 hours. The *E. coli* culture was then aliquoted, and used as a standard stock for all experiments. The *E. coli* stock was diluted ten-fold in PBS and spread on LB-ampicillin plates. After overnight incubation at 37°C, single colonies of *E. coli* were counted and the number was used to calculate the original concentration of the *E. coli* stock.

Surface chemistry

Two different surface chemistry methods were evaluated to immobilize antibodies for the highest capture efficiency of *E. coli* on chip (Figure 2B). Protein G is an immunoglobulin-binding protein and has the ability to immobilize the fragment crystallization (Fc) region of antibodies. NeutrAvidin has strong affinity for biotin-conjugated biomolecules (eg, biotinylated antibodies). In the first method, a biotinylated antibody was immobilized via NeutrAvidin-based surface chemistry.¹⁸ Briefly, a glass slide was first plasma treated to make the surface more hydrophilic. 200 mM of 3-mercaptopropyl-trimethoxysilane (100 μ L) dissolved in ethanol was then injected through the channels and incubated for 30 minutes at room temperature. 2 mM of N-(gamma-maleimidobutyryloxy) succinimide ester (GMBS; 100 μ L) in ethanol, a cross linker of proteins to antibodies, was then incubated in microchannels for 35 minutes at ambient temperature. An ethanol wash of 300 μ L was performed to remove the excess of untreated reagents after each incubation step. Before NeutrAvidin binding on the microchannels, 300 μ L of PBS was used to wash out ethanol from the chip surface. 30 μ L of NeutrAvidin solution (1 mg/mL in PBS) was incubated for 1 hour at room temperature before capture antibody immobilization. Three different experimental designs were performed for the NeutrAvidin studies. In the first design, 30 μ L of biotinylated lipopolysaccharide

(LPS) binding protein (LBP) Ab solution (5 μ g/mL), (goat immunoglobulin G antibody, R&D Systems, Minneapolis, MN) was subsequently immobilized on the microchannel surface. Then, 30 μ L of carrier-free recombinant human LBP (R&D Systems) at a concentration of 10 μ g/mL was coated on the top of biotinylated anti-LBP antibody (anti-LBP-LBP). In the second design, the same process as performed in the first design was followed by the addition of 30 μ L of bovine serum albumin (BSA) to block any remaining reactive spaces on the surface (anti-LBP-LBP-BSA). In the third design, first LBP was directly bound to GMBS to test the effect of protein orientation on *E. coli* capture, and then the BSA blocking step was performed after LBP binding (LBP-BSA). LBP was incubated in microchannels for 1 hour at room temperature.

In the second surface chemistry method, three nonbiotinylated antibodies (ie, anti-LPS antibody, cluster of differentiation 14 protein (CD14) of human monocyte and anti-flagellin antibody) were immobilized onto microchannels via Protein G.²⁶ In this method, processing steps were same as the first method except for the use of Protein G instead of NeutrAvidin. The Protein G concentration used in the immobilization of alternate antibodies onto the surface was 1 mg/mL. Then 30 μ L of 5 μ g/mL anti-LPS antibody (Abcam, Cambridge, MA) was immobilized on the surface via Protein G. 30 μ L of CD14 (Abcam, Cambridge, MA) was used at 5 μ g/mL to capture *E. coli* by binding to LPS on *E. coli* surface. Movement of LPS monomers to a binding site on CD14 was catalyzed by LBP.^{27,28} Finally, 30 μ L of anti-flagellin antibody (BioLegend, San Diego, CA) was used at a concentration of 5 μ g/mL to capture *E. coli* via the surface protein flagellin (part of flagellum).²⁹ These three capture agents were simultaneously incubated in microchannels for 1 hour at room temperature.

Sample preparation

EmGFP-expressing *E. coli* was spiked into 1 \times PBS, blood, milk, and spinach samples with the final concentrations ranging from 50 to 4,000 colony forming units (CFUs) per milliliter for analysis on chip. Discarded de-identified whole blood (purchased from Research Blood Components, LLC, Cambridge, MA) was used in this study. Whole blood was spiked with *E. coli* and then inverted gently for 1 hour to enable homogenous distribution of *E. coli*. Spinach and milk samples were obtained from a local grocery store. Spinach leaves were washed and then spiked with *E. coli* before thorough mixing in a blender with 100 mL of sterile deionized H₂O. The resultant mixture containing *E. coli* was

filtered through a cell strainer (70 μM) to remove the residues of spinach leaves prior to microchip testing. Milk was spiked with *E. coli* and vortexed for 1 minute before testing.

Device operation and on-chip detection

To optimize the capture efficiency, 75 μL of EmGFP-expressing *E. coli* sample was pipetted into the functionalized microchannels, and then incubated at ambient temperature for 30 minutes. Following the incubation, microchannels were washed with 300 μL of PBS at a flow rate of 2 $\mu\text{L}/\text{minute}$ using a syringe pump (Harvard Apparatus, Holliston, MA) for 150 minutes. For further optimization of processing time 2, 5, and 10 $\mu\text{L}/\text{minute}$ wash was performed. After washing, captured *E. coli* was imaged using an inverted fluorescence microscope (Carl Zeiss, Thornwood, NY) through a GFP fluorescence filter (excitation wavelength 470 nm). For comparison, bright field and fluorescence images of *E. coli* taken under 10 \times and 100 \times magnification were merged (Figure 2C). The number of *E. coli* detected under 10 \times magnification using a GFP filter was counted manually.

Statistical analysis

To evaluate the effect of the capture agent on capture efficiency, we analyzed the experimental results ($n = 2-6$) using analysis of variance with Tukey's post-hoc test for multiple comparisons with statistical significance threshold set at 0.05 ($P < 0.05$). To evaluate the effect of flow rate on capture efficiency, we analyzed the experimental results ($n = 3-8$) using nonparametric Kruskal-Wallis test followed by pair-wise comparisons with nonparametric upper-tailed Mann-Whitney U test, with statistical significance threshold set at 0.05 ($P < 0.05$). When nonparametric pair-wise tests were performed, Bonferroni correction was used for multiple comparisons. The agreement between chip counts and the *E. coli* stock concentrations in different media (PBS, blood, milk, spinach) was evaluated by calculating Pearson product-moment correlation coefficient (r) with statistical significance threshold set at 0.05 ($P < 0.05$). The Bland-Altman comparison analysis was used to evaluate the repeatability of the chip counts using residual analysis in comparison to *E. coli* stock concentrations. The coefficient of repeatability was calculated as 1.96 times the standard deviations of the differences between chip measurements and *E. coli* stock concentrations. In the Bland-Altman analysis, a mean difference of zero indicates that the chip measurement is unbiased with respect to the standard. A clinically acceptable range indicates the interval within

which the difference would fall approximately 95% of the time. If the mean difference and the limit of agreement are within the clinically acceptable range, then the developed measurement method is deemed comparable to the standard. Statistical analyses were performed with Minitab (Release 14; Minitab Inc, State College, PA).

Results and discussion

We engineered the surface chemistry using immobilized antibodies, where the performance of the chip relies on nanoscale reactions on the microchannel surface. We evaluated antibodies specific to *E. coli* surface proteins using commonly reported antibody immobilization methods that provide a uniform distribution of antibodies on the capture surface in microfluidic channels, ie, Protein G and NeutrAvidin based methods. To develop a rapid microchip method for *E. coli* detection, four different capturing agents were immobilized in microchannels via these two surface chemistry methods. As shown in Figure 2B, anti-LPS, anti-flagellin, and CD14 were immobilized on the microchannel surface via Protein G and anti-LBP antibody was immobilized on the microchannel surface via NeutrAvidin. To investigate the binding of anti-LBP antibody, and to observe the effect of antibody orientation on *E. coli* capture. Three different experimental designs were performed. In NeutrAvidin experiments, capture efficiencies of anti-LBP-LBP, anti-LBP-LBP-BSA, and LBP-BSA were observed to be $71.8\% \pm 5\%$, $60.7\% \pm 2\%$, and $44.5\% \pm 5\%$, respectively (Figure 3A). In comparison, *E. coli* capture efficiencies of anti-flagellin, anti-LPS and CD14 obtained by using Protein G based surface chemistry, were $46.9\% \pm 3\%$, $41.5\% \pm 5\%$, and $41.0\% \pm 2\%$, respectively (Figure 3A). The capture efficiency via anti-LBP-LBP was observed to be significantly greater ($P < 0.05$) than the other capture agents that were used with Protein G based surface chemistry (Figure 3A). In our prior study, Protein G and NeutrAvidin exhibited similar efficiency to immobilize capture agents on microchannel surfaces.²⁶ Thus, the difference in capture efficiency of *E. coli* was mainly due to the affinity of capture agents, ie, anti-LPS, anti-flagellin, CD14, and LBP. The highest capture efficiency ($71.8\% \pm 5\%$) was observed in microchannels immobilized with anti-LBP in the presence of LBP. This observation is in accordance with a previous report, in which LBP was shown to bind *E. coli* with high affinity; anti-LBP antibody helped the protein to take favorable orientation for *E. coli* capture.²⁷ BSA blocking was used to prevent nonspecific binding in microchannels in addition to preventing the binding of anti-LBP antibodies and LBP onto the succinimide group




Figure 3 Comparison of the capture efficiency of *E. coli* by two different surface chemistries and different capturing agents. *E. coli* were incubated at room temperature for 1 hour. **(A)** Three different experimental designs (anti-LBP-LBP, anti-LBP-LBP-BSA, and LBP-BSA) were performed on NeutrAvidin based surface chemistry. Three different capture agents were immobilized via Protein G based surface chemistry. The wash flow rate was 2 $\mu\text{L}/\text{min}$. Brackets connecting individual groups indicate statistically significant difference (analysis of Variance with Tukey's post-hoc test for multiple comparisons, $n = 2-6$, $P < 0.05$). **(B)** Effect of channel flow rate on capture efficiency of *E. coli* on chip. 75 μL of *E. coli* was flowed into microchannels. After sample incubation for 15 min at ambient temperature, three different wash flow rates (2, 5, and 10 $\mu\text{L}/\text{min}$) were used to optimize the capture efficiency of *E. coli* on chip. Statistical analysis indicated that flow rate had a significant effect on capture efficiency (nonparametric Kruskal–Wallis test), where 2 $\mu\text{L}/\text{min}$ resulted in significantly greater ($P < 0.05$) capture efficiency compared to 10 $\mu\text{L}/\text{min}$ flow rate. Brackets connecting individual groups indicate statistically significant difference. Data are presented as average \pm SEM. Non-parametric upper-tailed Mann–Whitney U test for pair-wise comparisons, $n = 3-8$, $P < 0.05$.

Abbreviations: BSA, bovine serum albumin; *E. coli*, *Escherichia coli*; LBP, lipopolysaccharide binding protein; SEM, standard error of the mean.

of GMBS. The use of BSA as a blocking agent did not result in a statistically significant difference in capture efficiency with both anti-LBP-LBP and anti-LBP-LBP-BSA. The LBP-BSA experiment showed that the orientation of the LBP protein is critical to capture of *E. coli*, which was supported by statistical analysis of experimental results as shown in Figure 3A. As an overall result, NeutrAvidin-mediated anti-LBP antibody-LBP immobilization performed on the microchannel surface to attain high capture efficiency of *E. coli*.

By investigating the effects of washing flow rate on *E. coli* capture efficiency, anti-LBP antibody based surface chemistry was further optimized. At the flow rates of 2, 5, and 10 $\mu\text{L}/\text{min}$, capture efficiencies were 70.7% \pm 4%, 60.5% \pm 3%, and 53.9% \pm 8%, respectively (Figure 3B).

Statistical analysis on experimental results indicated that flow rate had a significant effect on capture efficiency (nonparametric Kruskal–Wallis test), where 2 $\mu\text{L}/\text{min}$ resulted in significantly greater ($P < 0.05$) capture efficiency compared to 10 $\mu\text{L}/\text{min}$. The lower efficiency observed at higher flow rates may be related to the correspondingly higher shear stress within the microchannels. Additionally, we used food dyes to visualize and qualitatively analyze the wash steps in microchannels. We observed that the selected flow rate (2 $\mu\text{L}/\text{min}$) achieved effective removal of food dye solution from microchannels (Figure 4).

To determine the microchip's limit of detection for *E. coli* capture, we used LB agar plate culture as the gold standard for *E. coli* detection. We correlated agar plate results for a




Figure 4 **(A)** Three different food dye solutions were injected into microchannels before performing wash steps. **(B)** Images of channels before and after wash steps indicated that food dye was removed from microchannels at a flow rate of 2 $\mu\text{L}/\text{minute}$.




Figure 5 Correlation of *E. coli* quantification by microchip and LB plating. This experiment was performed to establish the correlation between bacteria cell counts obtained by colony count from LB agar plates and cell count after capture on a microfluidic device (**A**, **C**, **E** and **G**). Bland-Altman analysis between the microchip count and *E. coli* stock concentrations did not display an evidence for a systematic bias for chip counts. (**A**) 75 μ L of varying concentrations (up to 500 CFUs/mL) of *E. coli* spiked in PBS was injected into microchannels functionalized with anti-LBP antibody. For comparison, 75 μ L of each concentration of *E. coli* was plated out on ampicillin containing LB agar plates and incubated overnight. The number of *E. coli* colonies was counted the next day and compared to the *E. coli* counted on chip. The detection limit of microchip was found as 50 CFUs/mL. Data are presented as average \pm SEM ($n = 3$) ($r = 0.960$, $P = 0.009$). (**B**) The mean bias for *E. coli* spiked in PBS was -70 CFUs/mL sample in microchip counts compared to *E. coli* stock concentrations. (**C**) Varying concentrations (up to 400 CFUs/mL) of *E. coli* spiked in blood were injected into microchannels functionalized with anti-LBP antibody and the detection limit of microchip was found as 50 CFUs/mL ($r = 0.989$, $P = 0.011$). (**D**) The mean bias was -165 CFUs/mL of blood in microchip counts compared to *E. coli* stock concentrations. (**E**) Varying concentrations (up to 400 CFUs/mL) of *E. coli* spiked in milk were injected into microchannels functionalized with anti-LBP antibody and the detection limit of microchip was found as 50 CFUs/mL ($r = 0.962$, $P = 0.038$). (**F**) The mean bias was -163 CFUs/mL of milk in microchip counts compared to *E. coli* stock concentrations. (**G**) Varying concentrations (up to 4,000 CFUs/mL) of *E. coli* spiked in spinach were injected into microchannels functionalized with anti-LBP antibody and the detection limit of microchip was found as 500 CFUs/mL ($r = 0.977$, $P = 0.023$). (**H**) The mean bias was -1869 CFUs/mL of spinach sample in microchip counts compared to *E. coli* stock concentrations. (“ r ” indicates Pearson product-moment correlation coefficient, “ P ” indicates the statistical significance of correlation).

Abbreviations: CFU, colony forming unit; *E. coli*, *Escherichia coli*; LB, Luria-Bertani; LBP, lipopolysaccharide binding protein; SD, standard deviation; SEM, standard error of the mean.

series of known concentrations of *E. coli* with the microchip capture results. *E. coli* concentration and microchip-based *E. coli* detection showed a positive correlation for *E. coli* concentrations ranging from 50 to 4,000 CFUs/mL (Figure 5). The linearity of the correlation indicates that microchip-based *E. coli* detection can be used as an alternative to agar plate culture.

Figure 5A shows a sensitivity of 50 CFUs/mL that was obtained by the microchip count in PBS. Various sample types including blood, milk, and spinach, were used for rapid on-chip detection of *E. coli* (Figure 5C, E, and G). The detection limits of *E. coli* in blood, milk, and spinach samples were observed to be 50, 50, and 500 CFUs/mL as compared to *E. coli* dilution count, respectively. According to Bland–Altman analysis, the mean biases were -70 CFUs/mL, -165 CFUs/mL, -163 CFUs/mL, and -1869 CFUs/mL for PBS (50 to 500 CFUs/mL), blood (50 to 500 CFUs/mL), milk (50 to 500 CFUs/mL), and spinach (500 to 4,000 CFUs/mL) samples, respectively (Figure 5B, D, F, and H). Compared to *E. coli* spiked in PBS, the capture efficiency and limit of detection slightly decreased in other samples due to the immunological and enzymatic components in blood, milk, and spinach. For instance, there is a significant amount of albumin present in blood; this is commonly used as a blocking agent in immunoassays, and as such may have had a similar effect on our experiment. Milk also tends to have excessive albumin, casein, and other proteins, which may cause further blocking of the antigen-antibody binding. As the data show, these blocking and enzymatic agents can affect the capture efficiency. *E. coli* detection in spinach samples showed higher limit-of-detection than milk and blood experiments, potentially due to this experiment requiring preprocessing steps to remove the residuals of spinach leaves prior to microchip testing. Preprocessing steps (eg, filtering) caused a loss in *E. coli* cells before detection. On the other hand, the presented microchip assay achieved a lower detection limit for similar assays than was previously reported by methods such as magnetic separation method, fluorescence staining, and electrical detection (10^4 CFUs/mL and 8×10^4 CFUs/mL, respectively).^{30,31} Since nucleic acid-based-detection methods require more extensive sample preprocessing, antibody based *E. coli* detection on a chip presents a simpler alternative method to nucleic acid-based amplification assays on a chip.^{32,33} The current culture-based systems are subject to giving false negatives since only a small sample can be cultured compared to the whole sample volume, eg, 0.1% of the total blood volume is cultured in sepsis patients. This leads to 7.2 to 21.2%

of such cultures providing false negatives.^{10,11} With the microchip system, larger sample sizes can be processed by continuous flow, giving results much faster than a culture system. The sample can also be reprocessed within either the same or new microchannels; this would be expected to increase the number of pathogens captured from a single sample. However, for detection purposes, this would be not necessary granted that accurate detection results are provided by the chip. Capturing larger numbers of pathogens can be beneficial for other applications such as culturing and downstream genomic analysis.

In this study, we used *E. coli* as a model pathogen because *E. coli* have been extensively characterized, and as such it provides simple comparison to existing systems. We transformed BL21 Star™ with EmGFP-expressing plasmid and developed a microfluidic device to rapidly capture and quantify *E. coli* from various biological samples. The GFP-expressing *E. coli* was used to facilitate the quantification steps. The bacteria can be visualized either under a microscope in a laboratory setting, or for rapid counts in POC settings, with wide field of view lensless imaging systems for rapid counts for POC settings.^{20,24,25,34} Also, for characterization of *E. coli* capture from various biological samples, we employed a GFP-expressing strain for ease of detection under a fluorescence microscope. The extension from a GFP-expressing strain to wild-type strains can be simply achieved using an ELISA-based detection method, as previously reported.³⁵ The characterized microchip can be coupled with POC detection technologies such as surface plasmon resonance,^{36–38} on-chip ELISA,³⁹ and lensless imaging^{20,24} to achieve POC testing without reference to laboratory equipment. Additionally, the presented microfluidic approach can be adapted to detect other microorganisms that cause sepsis such as Gram-negative bacteria (*E. coli*, *Klebsiella*, *Pseudomonas*, *Enterobacter*, and *Hemophilus* species) and Gram-positive bacteria (*Staphylococcus*, *Streptococcus*, and *Enterococcus* species).⁴⁰ Also, the current material cost, without the labor and other costs of the microchip, includes 10¢ of glass, 1¢ of PMMA, and 70¢ of antibodies. The antibody cost could be lowered with large-scale production and ordering.

In conclusion, we presented a microfluidic immunoassay to capture *E. coli* from blood, milk and spinach samples. This microchip enables rapid detection of bacteria in blood samples and food supply. The technology can be potentially adapted for on-site real-time food quality monitoring and for diagnosis of sepsis at the POC.

Acknowledgments

This work was made possible by a research grant that was awarded and administered by the US Army Medical Research and Materiel Command and the Telemedicine and Advanced Technology Research Center, at Fort Detrick, MD. We thank Drs Paul Nisson and John Carney for their feedback and productive discussions on the project. We would also like to thank Mr Noah MacCallum for his feedback.

Disclosure

The authors report no conflicts of interest in this work.

References

- Schlichting D, McCollam JS. Recognizing and managing severe sepsis: a common and deadly threat. *South Med J*. 2007;100(6):594–600.
- Rangel JM, Sparling PH, Crowe C, Griffin PM, Swerdlow DL. Epidemiology of *Escherichia coli* O157:H7 outbreaks, United States, 1982–2002. *Emerg Infect Dis*. 2005;11(4):603–609.
- Crutchfield S, Roberts T. FoodReview: The 1990's: a dynamic decade for the U.S. food system. *Food Review Archives*. 2000;23(3).
- Frank C, Werber D, Cramer JP, et al. Epidemic profile of Shiga-toxin-producing *Escherichia coli* O104:H4 outbreak in Germany. *N Engl J Med*. 2011;365(19):1771–1780.
- Martin GS, Mannino DM, Eaton S, Moss M. The epidemiology of sepsis in the United States from 1979 through 2000. *N Engl J Med*. 2003;348(16):1546–1554.
- Wisplinghoff H, Bischoff T, Tallent SM, Seifert H, Wenzel RP, Edmond MB. Nosocomial bloodstream infections in US hospitals: analysis of 24,179 cases from a prospective nationwide surveillance study. *Clin Infect Dis*. 2004;39(3):309–317.
- Zhao CW, Ge BL, De Villena J, et al. Prevalence of *Campylobacter* spp, *Escherichia coli*, and *Salmonella* serovars in retail chicken, turkey, pork, and beef from the Greater Washington, DC, area. *Appl Environ Microbiol*. 2001;67(12):5431–5436.
- Rodgers GL, Mortensen J, Fisher MC, et al. Predictors of infectious complications after burn injuries in children. *Pediatr Infect Dis J*. 2000;19(10):990–995.
- Keen EF III, Robinson BJ, Hospenthal DR, et al. Incidence and bacteriology of burn infections at a military burn center. *Burns*. 2010;36(4):461–468.
- Fan ST, Teoh-Chan CH, Lau KF. Evaluation of central venous catheter sepsis by differential quantitative blood culture. *Eur J Clin Microbiol Infect Dis*. 1989;8(2):142–144.
- Quilici N, Audibert G, Conroy MC, et al. Differential quantitative blood cultures in the diagnosis of catheter-related sepsis in intensive care units. *Clin Infect Dis*. 1997;25(5):1066–1070.
- Mukkur TK, Bradley RE. Differentiation of avian thrombocytes from leukocytes by use of Giemsa's stain. *Poult Sci*. 1967;46(6):1595–1596.
- Sanderson MW, Sreerama S, Nagaraja TG. Sensitivity of direct plating for detection of high levels of *E. coli* O157:H7 in bovine fecal samples. *Curr Microbiol*. 2007;55(2):158–161.
- Jordan JA, Durso MB. Real-time polymerase chain reaction for detecting bacterial DNA directly from blood of neonates being evaluated for sepsis. *J Mol Diagn*. 2005;7(5):575–581.
- Wang S, Xu F, Demirci U. Advances in developing HIV-1 viral load assays for resource-limited settings. *Biotechnol Adv*. 2010;28(6):770–781.
- Cheng X, Irimia D, Dixon M, et al. A microfluidic device for practical label-free CD4(+) T cell counting of HIV-infected subjects. *Lab Chip*. 2007;7(2):170–178.
- Cheng X, Irimia D, Dixon M, et al. A microchip approach for practical label-free CD4+ T-cell counting of HIV-infected subjects in resource-poor settings. *J Acquir Immune Defic Syndr*. 2007;45(3):257–261.
- Kim YG, Moon S, Kuritzkes DR, Demirci U. Quantum dot-based HIV capture and imaging in a microfluidic channel. *Biosens Bioelectron*. 2009;25(1):253–258.
- Lee WG, Kim YG, Chung BG, Demirci U, Khademhosseini A. Nano/microfluidics for diagnosis of infectious diseases in developing countries. *Adv Drug Deliv Rev*. 2010;62(4–5):449–457.
- Moon S, Keles HO, Ozcan A, et al. Integrating microfluidics and lensless imaging for point-of-care testing. *Biosens Bioelectron*. 2009;24(11):3208–3214.
- Mahalanabis M, Al-Muayad H, Kulinski MD, Altman D, Klapperich CM. Cell lysis and DNA extraction of gram-positive gram-negative bacteria from whole blood in a disposable microfluidic chip. *Lab Chip*. 2009;9:2811–2817.
- Bennett AR, MacPhee S, Betts RP. The isolation and detection of *Escherichia coli* O157 by use of immunomagnetic separation and immunoassay procedures. *Lett Appl Microbiol*. 1996;22(3):237–243.
- Onoue Y, Konuma H, Nakagawa H, Hara-Kudo Y, Fujita T, Kumagai S. Collaborative evaluation of detection methods for *Escherichia coli* O157:H7 from radish sprouts and ground beef. *Int J Food Microbiol*. 1999;46(1):27–36.
- Moon S, Gurkan UA, Blander J, et al. Enumeration of CD4 T-Cells Using a Portable Microchip Count Platform in Tanzanian HIV-Infected Patients. *PLoS One*. 2011;6(7):e21409.
- Alyassin MA, Moon S, Keles HO, et al. Rapid automated cell quantification on HIV microfluidic devices. *Lab Chip*. 2009;9(23):3364–3369.
- Wang SQ, Esfahani M, Gurkan UA, Inci F, Kuritzkes DR, Demirci U. Efficient on-chip isolation of HIV subtypes. *Lab Chip*. 2012;12(8):1508–1515.
- Hailman E, Lichenstein HS, Wurfel MM, et al. Lipopolysaccharide (LPS)-binding protein accelerates the binding of LPS to CD14. *J Exp Med*. 1994;179(1):269–277.
- Yu B, Hailman E, Wright SD. Lipopolysaccharide binding protein and soluble CD14 catalyze exchange of phospholipids. *J Clin Invest*. 1997;99(2):315–324.
- Molloy MP, Herbert BR, Slade MB, et al. Proteomic analysis of the *Escherichia coli* outer membrane. *Eur J Biochem*. 2000;267(10):2871–2881.
- Yang L, Li Y. Simultaneous detection of *Escherichia coli* O157:H7 and *Salmonella* Typhimurium using quantum dots as fluorescence labels. *Analyst*. 2006;131(3):394–401.
- Varshney M, Li YB, Srinivasan B, Tung S. A label-free, microfluidics and interdigitated array microelectrode-based impedance biosensor in combination with nanoparticles immunoseparation for detection of *Escherichia coli* O157:H7 in food samples. *Sens Actuators B Chem*. 2007;128(1):99–107.
- Mahalanabis M, Do J, Almuayad H, Zhang JY, Klapperich CM. An integrated disposable device for DNA extraction and helicase dependent amplification. *Biomed Microdevices*. 2010;12(2):353–359.
- Min J, Kim JH, Lee Y, et al. Functional integration of DNA purification and concentration into a real time micro-PCR chip. *Lab Chip*. 2011;11(2):259–265.
- Gurkan UA, Moon S, Geckil H, et al. Miniaturized lensless imaging systems for cell and microorganism visualization in point-of-care testing. *Biotechnol J*. 2011;6(2):138–149.
- Nekhotiaeva N, Awasthi SK, Nielsen PE, Good L. Inhibition of *Staphylococcus aureus* gene expression and growth using antisense peptide nucleic acids. *Mol Ther*. 2004;10(4):652–659.
- Fernandez F, Hegnerova K, Pilarik M, Sanchez-Baeza F, Homola J, Marco MP. A label-free and portable multichannel surface plasmon resonance immunosensor for on site analysis of antibiotics in milk samples. *Biosens Bioelectron*. 2010;26(4):1231–1238.
- Chinowsky TM, Soelberg SD, Baker P, et al. Portable 24-analyte surface plasmon resonance instruments for rapid, versatile biodetection. *Biosens Bioelectron*. 2007;22(9–10):2268–2275.

38. Turker B, Guner H, Ayas S, et al. Grating coupler integrated photodiodes for plasmon resonance based sensing. *Lab Chip*. 2011;11(2):282–287.
39. Wang S, Zhao X, Khimji I, et al. Integration of cell phone imaging with microchip ELISA to detect ovarian cancer HE4 biomarker in urine at the point-of-care. *Lab Chip*. 2011;11(20):3411–3418.
40. Mahalanabis M, Al-Muayad H, Kulinski MD, Altman D, Klapperich CM. Cell lysis and DNA extraction of gram-positive and gram-negative bacteria from whole blood in a disposable microfluidic chip. *Lab Chip*. 2009;9(19):2811–2817. Epub December 8, 2009. doi: 10.1039/b905065p. PubMed PMID: 19967118.

International Journal of Nanomedicine

Dovepress

Publish your work in this journal

The International Journal of Nanomedicine is an international, peer-reviewed journal focusing on the application of nanotechnology in diagnostics, therapeutics, and drug delivery systems throughout the biomedical field. This journal is indexed on PubMed Central, MedLine, CAS, SciSearch®, Current Contents®/Clinical Medicine,

Journal Citation Reports/Science Edition, EMBase, Scopus and the Elsevier Bibliographic databases. The manuscript management system is completely online and includes a very quick and fair peer-review system, which is all easy to use. Visit <http://www.dovepress.com/testimonials.php> to read real quotes from published authors.

Submit your manuscript here: <http://www.dovepress.com/international-journal-of-nanomedicine-journal>